

Report for:

Mr. Quality Control
EMLab P&K (QA)
6000 Shoreline Ct
Ste 205
South San Francisco, CA 94080

Regarding: Eurofins EPK Built Environment Testing, LLC
Project: Sample Report
EML ID: 2404396

Approved by:

Dates of Analysis:
PCR - Coronavirus (CoV) Screen: 05-12-2020



Technical Manager
Ariunaa Jalsrai

Service SOPs: PCR - Coronavirus (CoV) Screen (EM-BC-S-8576)

All samples were received in acceptable condition unless noted in the Report Comments portion in the body of the report. Due to the nature of the analyses performed, field blank correction of results is not applied. The results relate only to the samples as received and tested.

Eurofins EPK Built Environment Testing, LLC ("the Company"), a member of the Eurofins Built Environment Testing group of companies, shall have no liability to the client or the client's customer with respect to decisions or recommendations made, actions taken or courses of conduct implemented by either the client or the client's customer as a result of or based upon the Test Results. In no event shall the Company be liable to the client with respect to the Test Results except for the Company's own willful misconduct or gross negligence nor shall the Company be liable for incidental or consequential damages or lost profits or revenues to the fullest extent such liability may be disclaimed by law, even if the Company has been advised of the possibility of such damages, lost profits or lost revenues. In no event shall the Company's liability with respect to the Test Results exceed the amount paid to the Company by the client therefor.

Client: EMLab P&K (QA)
 C/O: Mr. Quality Control
 Re: Sample Report

Date of Sampling: 05-12-2020
 Date of Receipt: 05-12-2020
 Date of Report: 05-12-2020

Coronavirus (CoV) detection by RT RealTime PCR

Location:	#1	#2
Comments (see below)	None	None
Lab ID-Version‡:	11466463-1	11466464-1
Sample Size:	1	1
Sample Type:	Swab sample	Swab sample

Test Results:

Coronavirus (CoV)	Presumptive	Not Detected
Presumptive = Coronavirus RNA was detected; however specific identification and confirmation of SARS-CoV-2 (COVID-19) RNA requires further testing.		
Coronavirus (CoV) constitutes a family of viruses which have a protein envelope. The VIRSeek SARSCoV-2 Screen assay developed by Eurofins detects the gene that codes for that envelope. These include SARS-CoV, MERS-CoV, SARS-CoV-2 (COVID-19), etc.		

Comments:

Sampling techniques, contaminants infecting samples, unrepresentative samples and other similar or dissimilar factors may affect results. Eurofins EMLab P&K hereby disclaims any liability for indirect, punitive, incidental, special or consequential damages arising out of the use or interpretation of the data contained in, or any actions taken in reliance upon, this report. Any actual direct damages arising out of the use or interpretation of the data contained in, or any actions or omitted taken in reliance upon, this report shall be limited to the cost of this report. The testing reflected in this report has been validated by using a surrogate virus (Murine Norovirus) at concentrations of approximately 10⁴ TCID₅₀. Virus concentrations below the tested parameters may not be sufficient for positive detection.

‡ A "Version" indicated by "-x" after the Lab ID# with a value greater than 1 indicates a sample with amended data. The revision number is reflected by the value of "x".