

Client: EMLab P&K (QA)
 C/O: Mr. Quality Control
 Re: Sample Report

Date of Sampling: 02-01-2013
 Date of Receipt: 02-01-2013
 Date of Report: 02-08-2013

CULTURABLE AIR FUNGI REPORT

Location:	1: Outside Reference		2		3	
Comments (see below)	None		None		None	
Lab ID-Version‡:	4590863-1		4590864-1		4590865-1	
Analysis Date:	02/08/2013		02/08/2013		02/08/2013	
Medium:	MEA		MEA		MEA	
	raw ct.	cfu*/m3	raw ct.	cfu*/m3	raw ct.	cfu*/m3
Acremonium					1	12
Alternaria	1	12				
Aspergillus flavus						
Aspergillus fumigatus						
Aspergillus nidulans						
Aspergillus niger	3	35	1	12		
Aspergillus ochraceus						
Aspergillus terreus					2	24
Aspergillus versicolor	1	12				
Aureobasidium						
Basidiomycetes	1	12				
Bipolaris/Drechslera group						
Botrytis						
Chaetomium						
Cladosporium	26	320	13	150	9	110
Curvularia						
Epicoccum						
Fusarium						
Non-sporulating fungi	5	59	1	12		
Paecilomyces						
Penicillium	2	24	5	59	36	450
Phoma						
Rhizopus						
Stachybotrys chartarum						
Ulocladium						
Yeasts			1	12		
Positive Hole	400		400		400	
Sample volume (liters)	85		85		85	
§ TOTAL CFU*/M3		470		250		590

* cfu = colony forming units Positive hole correction chart used for all calculations

Comments:

Note: Interpretation is left to the company and/or persons who conducted the field work.

‡ A "Version" indicated by "-x" after the Lab ID# with a value greater than 1 indicates a sample with amended data. The revision number is reflected by the value of "x".

§ Total CFU/m3 has been rounded to two significant figures to reflect analytical precision.

Fungal culture types listed without a count or data entry were not detected during the course of the analysis for the respective sample.