


Report for:

Mr. Quality Control
EMLab P&K (QA)
1150 Bayhill Drive
Suite 100
San Bruno, CA 94066

Regarding: Eurofins EPK Built Environment Testing, LLC
Project: Sample Report
EML ID: 1052830

Approved by:

Dates of Analysis:
USP 797-Media Fill Test: 04-19-2013



Technical Manager
Murali Putty

Service SOPs: USP 797-Media Fill Test (EM-USP-S-2083)

All samples were received in acceptable condition unless noted in the Report Comments portion in the body of the report. Due to the nature of the analyses performed, field blank correction of results is not applied. The results relate only to the samples as received and tested.

Eurofins EPK Built Environment Testing, LLC ("the Company"), a member of the Eurofins Built Environment Testing group of companies, shall have no liability to the client or the client's customer with respect to decisions or recommendations made, actions taken or courses of conduct implemented by either the client or the client's customer as a result of or based upon the Test Results. In no event shall the Company be liable to the client with respect to the Test Results except for the Company's own willful misconduct or gross negligence nor shall the Company be liable for incidental or consequential damages or lost profits or revenues to the fullest extent such liability may be disclaimed by law, even if the Company has been advised of the possibility of such damages, lost profits or lost revenues. In no event shall the Company's liability with respect to the Test Results exceed the amount paid to the Company by the client therefor.

Client: EMLab P&K (QA)
C/O: Mr. Quality Control
Re: Sample Report

Date of Sampling: 04-05-2013
Date of Receipt: 04-05-2013
Date of Report: 04-19-2013

USP <797> Aseptic Manipulation - Media-Fill Test

Location:	3	4	5
Comments (see below)	None	None	None
Sample type	Media-filled vial	Media-filled vial	Media-filled vial
Media used	TSB	TSB	TSB
Lab ID-Version‡:	4734356-1	4734357-1	4734358-1
Visible Turbidity*	Positive	Negative	Negative

*Failure of a media fill test is indicated by visible turbidity (Positive) in any one of the media-fill units on or before 14 days.†

Comments:

†The United States Pharmacopeial Convention. <797> Pharmaceutical Compounding - Sterile Preparations. *Revision Bulletin*. 2008, p. 12.

‡ A "Version" indicated by -"x" after the Lab ID# with a value greater than 1 indicates a sample with amended data. The revision number is reflected by the value of "x".


Client: EMLab P&K (QA)
C/O: Mr. Quality Control
Re: Sample Report

Date of Sampling: 04-05-2013
Date of Receipt: 04-05-2013
Date of Report: 04-19-2013

USP <797> Aseptic Manipulation - Media-Fill Test

PROJECT ANALYST AND SIGNATORY REPORT

Project Analyst



Analyst: Malcolm Moody

The test report shall not be reproduced except in full, without written approval of the laboratory. The report must not be used by the client to claim product certification, approval, or endorsement by any agency of the federal government. The Company reserves the right to dispose of all samples after a period of thirty (30) days, according to all state and federal guidelines, unless otherwise specified.

‡ A "Version" indicated by "-x" after the Lab ID# with a value greater than 1 indicates a sample with amended data. The revision number is reflected by the value of "x".