

Report for:


Mr. Quality Control
EMLab P&K (QA)
1150 Bayhill Drive
Suite 100
San Bruno, CA 94066

Regarding: Eurofins EPK Built Environment Testing, LLC
Project: Sample Report
EML ID: 1225184

Approved by:

Dates of Analysis:

Quantitative spore count direct exam: 06-24-2014



Technical Manager
Murali Putty

Service SOPs: Quantitative spore count direct exam (EM-MY-S-1041)
AIHA LAP, LLC accredited service, Lab ID #102856

All samples were received in acceptable condition unless noted in the Report Comments portion in the body of the report. Due to the nature of the analyses performed, field blank correction of results is not applied. The results relate only to the samples as received and tested.

Eurofins EPK Built Environment Testing, LLC ("the Company"), a member of the Eurofins Built Environment Testing group of companies, shall have no liability to the client or the client's customer with respect to decisions or recommendations made, actions taken or courses of conduct implemented by either the client or the client's customer as a result of or based upon the Test Results. In no event shall the Company be liable to the client with respect to the Test Results except for the Company's own willful misconduct or gross negligence nor shall the Company be liable for incidental or consequential damages or lost profits or revenues to the fullest extent such liability may be disclaimed by law, even if the Company has been advised of the possibility of such damages, lost profits or lost revenues. In no event shall the Company's liability with respect to the Test Results exceed the amount paid to the Company by the client therefor.

Eurofins EPK Built Environment Testing, LLC's LabServe® reporting system includes automated fail-safes to ensure that all AIHA LAP, LLC quality requirements are met and notifications are added to reports when any quality steps remain pending.

Client: EMLab P&K (QA)
C/O: Mr. Quality Control
Re: Sample Report

Date of Sampling: 06-22-2014
Date of Receipt: 06-23-2014
Date of Report: 06-24-2014

QUANTITATIVE SPORE COUNT REPORT

Location:	1		2		3	
Comments (see below)	None		None		None	
Sample type	Bulk sample		Swab sample		Tape sample	
Lab ID-Version‡:	5574593-1		5574594-1		5574595-1	
Analysis Date:	06/24/2014		06/24/2014		06/24/2014	
Dilution	1:1		1:40		1:1	
	raw ct.	spores/unit	raw ct.	spores/unit	raw ct.	spores/unit
<i>Alternaria</i>			1	370		
Ascospores	2	43			2	2
Basidiospores	1	22	2	730	7	7
<i>Botrytis</i>						
<i>Chaetomium</i>			1	370		
<i>Cladosporium</i>	9	200	97	35,000	105	110
<i>Curvularia</i>						
<i>Epicoccum</i>						
<i>Fusarium</i>						
<i>Myrothecium</i>						
<i>Nigrospora</i>						
Other colorless						
<i>Penicillium/Aspergillus</i> types†	26	570	31	11,000	160	160
<i>Pithomyces</i>						
Rusts						
Smuts, Periconia, Myxomycetes			3	1,100		
<i>Stachybotrys</i>						
<i>Stemphylium</i>						
<i>Torula</i>						
<i>Ulocladium</i>						
<i>Zygomycetes</i>						
Background debris (1-4+)††	N/A		2+		2+	
Sample size	1		1		1	
Unit	1 cm2		1 cm2		1 cm2	
Hyphal fragments/unit	< 1		3,300		4	
§ TOTAL SPORES/UNIT		830		49,000		270

Comments:

† The spores of *Aspergillus* and *Penicillium* (and others such as *Acremonium*, *Paecilomyces*) are small and round with very few distinguishing characteristics. They cannot be differentiated by non viable sampling methods. Also, some species with very small spores are easily missed, and may be undercounted.

†† Background debris is an indication of the amount of non-biological particulate matter present on the slide (dust in the air) and is graded from 1+ to 4+ with 4+ indicating the largest amounts. This background material is also an indication of visibility for the analyst and resultant difficulty reading the slide. For example, high background debris may obscure the small spores such as the *Penicillium/Aspergillus* group. Counts from areas with 4+ background debris should be regarded as minimal counts and may actually be higher than reported.

Where tape lifts are performed for bulk sample analysis, the unit reported is specific to the area of tape analyzed. Where swabs are performed, the unit reported is specific to the swab analyzed.

Due to the inherent inhomogeneity and variable recovery rates of surface fungal samples, all reported spore counts are extrapolated estimates. Counts of fewer than 400 spores or greater than 40,000 spores are considered imprecise and not reproducible due to limitations in the counting method and equipment sensitivity.

All samples were received in acceptable condition unless noted in the Report Comments portion in the body of the report. Due to the nature of the analyses performed, field blank correction of results is not applied. The results relate only to the samples as received and tested.

‡ A "Version" indicated by "-x" after the Lab ID# with a value greater than 1 indicates a sample with amended data. The revision number is reflected by the value of "x".

§ Total Spores/unit has been rounded to two significant figures to reflect analytical precision.

The limit of detection is 1 spore per area analyzed; Analytical Sensitivity is 1 spore per unit times the dilution factor.


Spore types listed without a count or data entry were not detected during the course of the analysis for the respective sample, indicating a raw count of <1 spore.

Client: EMLab P&K (QA)
C/O: Mr. Quality Control
Re: Sample Report

Date of Sampling: 06-22-2014
Date of Receipt: 06-23-2014
Date of Report: 06-24-2014

QUANTITATIVE SPORE COUNT REPORT
PROJECT ANALYST AND SIGNATORY REPORT

Project Analyst



Analyst: Malcolm Moody

The test report shall not be reproduced except in full, without written approval of the laboratory. The report must not be used by the client to claim product certification, approval, or endorsement by AIHA LAP, LLC, or any agency of the federal government. The Company reserves the right to dispose of all samples after a period of thirty (30) days, according to all state and federal guidelines, unless otherwise specified.

‡ A "Version" indicated by "-x" after the Lab ID# with a value greater than 1 indicates a sample with amended data. The revision number is reflected by the value of "x".