

Report for:

**Mr. Quality Control**  
**EMLab P&K (QA)**  
1150 Bayhill Drive  
Suite 100  
San Bruno, CA 94066


---

Regarding: Eurofins EPK Built Environment Testing, LLC  
Project: Sample Report  
EML ID: 1225184

Approved by:

Dates of Analysis:

Quantitative spore count direct exam: 06-24-2014



Technical Manager  
Murali Putty

Service SOPs: Quantitative spore count direct exam (EM-MY-S-1041)  
AIHA-LAP, LLC accredited service, Lab ID #102856

---

All samples were received in acceptable condition unless noted in the Report Comments portion in the body of the report. Due to the nature of the analyses performed, field blank correction of results is not applied. The results relate only to the samples as received and tested.

Eurofins EPK Built Environment Testing, LLC ("the Company"), a member of the Eurofins Built Environment Testing group of companies, shall have no liability to the client or the client's customer with respect to decisions or recommendations made, actions taken or courses of conduct implemented by either the client or the client's customer as a result of or based upon the Test Results. In no event shall the Company be liable to the client with respect to the Test Results except for the Company's own willful misconduct or gross negligence nor shall the Company be liable for incidental or consequential damages or lost profits or revenues to the fullest extent such liability may be disclaimed by law, even if the Company has been advised of the possibility of such damages, lost profits or lost revenues. In no event shall the Company's liability with respect to the Test Results exceed the amount paid to the Company by the client therefor.

Eurofins EPK Built Environment Testing, LLC's LabServe® reporting system includes automated fail-safes to ensure that all AIHA-LAP, LLC quality requirements are met and notifications are added to reports when any quality steps remain pending.

Client: EMLab P&K (QA)  
C/O: Mr. Quality Control  
Re: Sample Report

Date of Sampling: 06-22-2014  
Date of Receipt: 06-23-2014  
Date of Report: 06-24-2014

## QUANTITATIVE SPORE COUNT REPORT

|                                |             |             |        |             |             |        |
|--------------------------------|-------------|-------------|--------|-------------|-------------|--------|
| Location:                      | 1           |             |        | 2           |             |        |
| Comments (see below)           | None        |             |        | None        |             |        |
| Sample type                    | Bulk sample |             |        | Swab sample |             |        |
| Lab ID-Version†:               | 5574593-1   |             |        | 5574594-1   |             |        |
| Analysis Date:                 | 06/24/2014  |             |        | 06/24/2014  |             |        |
| Dilution                       | 1:1         |             |        | 1:40        |             |        |
|                                | raw ct.     | spores/unit | growth | raw ct.     | spores/unit | growth |
| Alternaria                     |             |             |        | 1           | 370         | -      |
| Ascospores                     | 2           | 43          | -      |             |             |        |
| Basidiospores                  | 1           | 22          | -      | 2           | 730         | -      |
| Bipolaris/Drechslera group     |             |             |        |             |             |        |
| Botrytis                       |             |             |        |             |             |        |
| Chaetomium                     |             |             |        | 1           | 370         | -      |
| Cladosporium                   | 9           | 200         | -      | 97          | 35,000      | < 1+   |
| Curvularia                     |             |             |        |             |             |        |
| Epicoccum                      |             |             |        |             |             |        |
| Fusarium                       |             |             |        |             |             |        |
| Myrothecium                    |             |             |        |             |             |        |
| Nigrospora                     |             |             |        |             |             |        |
| Other colorless                |             |             |        |             |             |        |
| Penicillium/Aspergillus types‡ | 26          | 570         | -      | 31          | 11,000      | -      |
| Pithomyces                     |             |             |        |             |             |        |
| Rusts                          |             |             |        |             |             |        |
| Smuts, Periconia, Myxomycetes  |             |             |        | 3           | 1,100       | -      |
| Stachybotrys                   |             |             |        |             |             |        |
| Stemphylium                    |             |             |        |             |             |        |
| Torula                         |             |             |        |             |             |        |
| Ulocladium                     |             |             |        |             |             |        |
| Zygomycetes                    |             |             |        |             |             |        |
| Background debris (1-4+)††     | N/A         |             |        | 2+          |             |        |
| Sample size                    | 1           |             |        | 1           |             |        |
| Unit                           | 1 cm2       |             |        | 1 cm2       |             |        |
| Hyphal fragments/unit          | < 1         |             |        | 3,300       |             |        |
| <b>§ TOTAL SPORES/UNIT</b>     |             | 830         |        |             | 49,000      |        |

### Comments:

Dash ("-") means not detected. Quantities of molds seen growing on the surface sampled are listed in the "growth" column and are graded < 1+ to 4+, with 4+ denoting the highest numbers. This method differentiates between spores from mold growth and spores present because of normal fallout.

† The spores of *Aspergillus* and *Penicillium* (and others such as *Acremonium*, *Paecilomyces*) are small and round with very few distinguishing characteristics. They cannot be differentiated by non viable sampling methods. Also, some species with very small spores are easily missed, and may be undercounted.

†† Background debris is an indication of the amount of non-biological particulate matter present on the slide (dust in the air) and is graded from 1+ to 4+ with 4+ indicating the largest amounts. This background material is also an indication of visibility for the analyst and resultant difficulty reading the slide. For example, high background debris may obscure the small spores such as the *Penicillium/Aspergillus* group. Counts from areas with 4+ background debris should be regarded as minimal counts and may actually be higher than reported.

Where tape lifts are performed for bulk sample analysis, the unit reported is specific to the area of tape analyzed.

‡ A "Version" indicated by "-x" after the Lab ID# with a value greater than 1 indicates a sample with amended data. The revision number is reflected by the value of "x".

§ Total Spores/unit has been rounded to two significant figures to reflect analytical precision.

The limit of detection is 1 spore per area analyzed; Analytical Sensitivity is 1 spore per unit times the dilution factor.

Spore types listed without a count or data entry were not detected during the course of the analysis for the respective sample, indicating a raw count of <1 spore.

Client: EMLab P&K (QA)  
C/O: Mr. Quality Control  
Re: Sample Report

Date of Sampling: 06-22-2014  
Date of Receipt: 06-23-2014  
Date of Report: 06-24-2014

## QUANTITATIVE SPORE COUNT REPORT

| Location:                      | 3                 |             |        |
|--------------------------------|-------------------|-------------|--------|
| Comments (see below)           | None              |             |        |
| Sample type                    | Tape sample       |             |        |
| Lab ID-Version†:               | 5574595-1         |             |        |
| Analysis Date:                 | 06/24/2014        |             |        |
| Dilution                       | 1:1               |             |        |
|                                | raw ct.           | spores/unit | growth |
| <i>Alternaria</i>              |                   |             |        |
| Ascospores                     | 2                 | 2           | -      |
| Basidiospores                  | 7                 | 7           | -      |
| Bipolaris/Drechslera group     |                   |             |        |
| Botrytis                       |                   |             |        |
| Chaetomium                     |                   |             |        |
| Cladosporium                   | 105               | 110         | -      |
| Curvularia                     |                   |             |        |
| Epicoccum                      |                   |             |        |
| Fusarium                       |                   |             |        |
| Myrothecium                    |                   |             |        |
| Nigrospora                     |                   |             |        |
| Other colorless                |                   |             |        |
| Penicillium/Aspergillus types‡ | 160               | 160         | -      |
| Pithomyces                     |                   |             |        |
| Rusts                          |                   |             |        |
| Smuts, Periconia, Myxomycetes  |                   |             |        |
| Stachybotrys                   |                   |             |        |
| Stemphylium                    |                   |             |        |
| Torula                         |                   |             |        |
| Ulocladium                     |                   |             |        |
| Zygomycetes                    |                   |             |        |
| Background debris (1-4+)††     | 2+                |             |        |
| Sample size                    | 1                 |             |        |
| Unit                           | 1 cm <sup>2</sup> |             |        |
| Hyphal fragments/unit          | 4                 |             |        |
| <b>§ TOTAL SPORES/UNIT</b>     |                   | <b>270</b>  |        |

### Comments:

Dash ("-") means not detected. Quantities of molds seen growing on the surface sampled are listed in the "growth" column and are graded < 1+ to 4+, with 4+ denoting the highest numbers. This method differentiates between spores from mold growth and spores present because of normal fallout.

† The spores of *Aspergillus* and *Penicillium* (and others such as *Acremonium*, *Paecilomyces*) are small and round with very few distinguishing characteristics. They cannot be differentiated by non viable sampling methods. Also, some species with very small spores are easily missed, and may be undercounted.

†† Background debris is an indication of the amount of non-biological particulate matter present on the slide (dust in the air) and is graded from 1+ to 4+ with 4+ indicating the largest amounts. This background material is also an indication of visibility for the analyst and resultant difficulty reading the slide. For example, high background debris may obscure the small spores such as the *Penicillium/Aspergillus* group. Counts from areas with 4+ background debris should be regarded as minimal counts and may actually be higher than reported.

Where tape lifts are performed for bulk sample analysis, the unit reported is specific to the area of tape analyzed.

‡ A "Version" indicated by "-x" after the Lab ID# with a value greater than 1 indicates a sample with amended data. The revision number is reflected by the value of "x".

§ Total Spores/unit has been rounded to two significant figures to reflect analytical precision.

The limit of detection is 1 spore per area analyzed; Analytical Sensitivity is 1 spore per unit times the dilution factor.

Spore types listed without a count or data entry were not detected during the course of the analysis for the respective sample, indicating a raw count of <1 spore.